



Government of India  
Ministry of Communications & IT

Dept. of Telecom., Wireless Planning and Coordination Wing  
Regional Licensing Office, IMS Campus Ghitorni, New Delhi-110030



File No. J-22022/811/2017-RLO-NR/201422

Date : 20/06/2017

22.06.2017


To,

Omrom Healthcare India Private Limited  
6th Floor, B Block Sewa Tower Plot No. 19, Sector- 18, Undyog Vihar, Gurgaon-122008

Subject : Issuance of new Equipment Type Approval (ETA) Certificate Model No. **HEM-7600T-E**  
Equipment Name **Blood Pressure Monitor**

Sir,

With reference to your application dated 23/05/2017 received on 30/05/2017 on the above subject, it is to forward herewith an ETA No. NR- 6002 dated 20/06/2017 as desired therein.

  
Anil Verma

Dy. Wireless Adviser

उप बेतार सलाहकार  
Deputy Wireless Advisor

दूर संचार विभाग  
Department of Telecommunications  
क्षेत्रीय अनुज्ञप्ति कार्यालय (उ० क्षेत्र)  
Regional Licensing Office (N.R.)  
नई दिल्ली-110030





Government of India

Ministry of Communications & IT

Dept. of Telecom., Wireless Planning and Coordination Wing  
Regional Licensing Office, IMS Campus Ghitorni, New Delhi-110030



File No. J-22022/811/2017-RLO-NR/201422

Date : 20/06/2017

22.06.2017

ETA Certificate No: NR-ETA/6002-RLO(NR)

Equipment Type Approval is hereby granted for under mentioned equipment for operation with following parameters/conditions:

**I. Details of Applicant and parameters of Equipment:**

|                          |  |
|--------------------------|--|
| 1. Name of the Applicant | Omron Healthcare India Private Limited     |
| 2. Equipment             | Blood Pressure Monitor                     |
| 3. Model No.             | HEM-7600T-E                                |
| 4. Manufactured by       | Omron Healthcare Co., Ltd., Japan          |
| 5. Frequency Range (MHz) | 2402-2480MHz                               |
| 6. Max. Output Power     | 1.02 dBm                                   |
| 7. Modulation            | GFSK                                       |
| 8. Remarks               | Maximum EIRP should not exceed 4W(36 dBm). |

**II. Conditions:**

1. This is a clearance from technical angle only. The use of the equipment is regulated as per the provision mentioned in the Gazette Notification No. G.S.R. 45(E) dated 28.01.2005 as amended from time to time.
2. This approval will not be valid in case of any change in the above parameters and not conforming to the notification No. given in Para 1.
3. Use of such equipment has been exempted from licensing requirement vide Gazette Notification given in Para 1, on Non-Interference, Non-Protection and Sharing (Non-Executive) basis.
4. Use of this equipment with any change in the above parameters and not conforming to the above Notifications will require a specific wireless license from his Ministry.
5. Use of such equipment is also subject to the applicability/fulfillment of the specific service license as required from the Central Government.
6. This is not an IMPORT LICENSE. For the import of these equipments, a separate "IMPORT LICENSE" is required from respective RLO of WPC wing as per procedures.

**Note:** - (i) The above ETA is specific for the RF band/Model/Equipment mentioned above only.

(ii) Record of all the equipments imported needs to be maintained and submitted to this Ministry as and when required.

Anil Verma

22.06.2017

Deputy Wireless Adviser

Deputy Wireless Advisor

दूर संचार विभाग

Department of Telecommunications

राष्ट्रीय संचारिता कार्यालय (उ० क्षेत्र),

Regional Licensing Office (N.R.)

नई दिल्ली-110030